

Risk assessment

Support rails and handrails - 01.02.2017 (rev. 10.08.2017 – Translation 5.10.2023)

Product type and ID:

[REGAL 20203](#)

[REGAL 20108](#)

[REGAL 20204](#)

[OVAL 20111](#)

[OVAL 20115](#)

[OVAL 20112](#)

[OVAL 20116](#)

[OVAL 20114](#)

[OVAL 20140](#)

[SOLO HANDRAILS](#)

[MODULAR HANDRAILS](#)

Manufacturer intended use / uses for the product / products listed above:

Use while permanently mounted to a wall or equivalent.

Name of manufacturer, registered trade name or trademark and street address:

[Väinö Korpinen Oy](#)

[Kalkkipellontie 6](#)

[02650 Espoo](#)

[Finland](#)

Risk assessment:

Basic property / risk	Reason for risk elimination	Methods for addressing the residual risk
Load capacity / structural failure	Fulfills EN 12182 (16.2) general requirements and test methods	-
Wall mounting / mounting failure	Mounting according to instructions	Mounting inspection

01.02.2017 (rev. 10.08.2017 – Translation 5.10.2023)

The manufacturer has implemented Factory Production Control (FPC) in accordance with the Construction Products Directive.

The risk assessment has applied EN ISO 14971 where applicable.